2025/12/01 01:01 1/3 Profiling, Hemodynamic

# **Profiling, Hemodynamic**

<mermaid>

```
flowchart TD
filter_dys("(1) Bad LVEF OR (2) low CI AND low HR`")
type_dys["Dysfunctional"]
dys_svr_high["consider vasodilators"]
dys_svr_low["Consider vasoconstrictors"]
dys_hypoperf["End organ hypoperfusion?"]
dys_hypoperf_yes["Access SVR"]
dys_hypoperf_no
filter_dys -- Yes --> type_dys
type_dys -- --> dys_hypoperf
dys_hypoperf -- Yes --> dys_hypoperf_yes
dys_hypoperf_yes -- SVR > 1200 --> dys_svr_high
dys_hypoperf_yes -- SVR < 800 AND MAP < 65 --> dys_svr_low
```

</mermaid>

# **Hepatic Congestion**

- Article, VExUS Grading
- LV Diastolic Function
- Echopedia, Diastolic Function
- Hepatic Vascular Pulsatility

#### FREE Exam

#### Setup

- Enter patient information
- Attach EKG leads
- Select phased array transducer
- Choose ST UMMC 1 Echo preset

#### Parasternal long axis (PLAX)

- Qualitative EF assessment (CLIP)
- LVOT diameter (SAVE)
- PLAX, during end systole
- Inner edge to inner edge of aortic at base of aortic valve
- Normal 1.8-2.4 (~BSA, can use as surrogate if unable to measure)
- Color doppler over MV and AoV to look for regurgitation (CLIP)

## Parasternal short axis (PSAX)

- Qualitative EF assessment at each level
- Level of papillary muscles (CLIP) assess RV as well
- Level of mitral valve (CLIP)
- Level of aortic valve (CLIP)
- Color doppler over tricuspid to check for TR (CLIP)

## Apical four chamber (A4C)

- Qualitative assessment of RV and LV size (CLIP)
- Color doppler over MV, LA, and LV (CLIP)
- Mitral inflow E-a (SAVE)
  - A4C, mitral valve, PW doppler just inside ventricle
  - Above baseline, measuring flow into the ventricle/towards the probe
  - E = early diastolic filling
  - ∘ A = late atrial kick
  - A is just before QRS, E is before A
  - E > A in normal and pseudonormal (super abnormal)
- Mitral annulus TDI (SAVE)
  - A4C, mitral valve, lateral annulus, TDI → PW
  - A' is just before QRS, E' is before A'
  - E' and E occur at the same time point in the cardiac cycle
  - Normal E/E' >10
- Color doppler over TV (CLIP)
- TR Vmax (SAVE)
  - A4C, CW doppler
  - Can also be done in PSAX, CW doppler, if visible at aortic valve level
  - Only if tricuspid regurgitation is present
  - Surrogate for RVSP/PASP (TR max PG = RVSP + CVP)
- TAPSE (SAVE)
  - A4C, tricuspid valve, lateral annulus, M-mode
  - Estimate visually before measuring
  - Measure peak to valley
  - RV specific, only free wall, no contribution from septum/LV
  - ∘ Normal >1.7

# Apical five chamber (A5C)

- Collar doppler over LVOT and AoV (CLIP)
- LVOT VTI (SAVE)
  - A5C, aortic valve, PW doppler where LVOT diameter was measured
  - Quantitative surrogate for stroke volume (SV)
  - Trace largest flow away from probe, baseline to baseline
  - Normal 18-24 in euvolemia (approx. 10x BSA)
- Stroke volume variation (SAVE)
  - Using doppler saved for LVOT VTI, need at least 10 beats
  - Decrease sweep speed (25-35 mm/sec) to see multiple beats

https://ewrobbins.com/ Printed on 2025/12/01 01:01

- Measure SV maximum and minimum flow
- Cannot do in arrhythmia, not validated in low EF
- AoV Vmax
  - A5C, aortic valve, CW doppler
  - If AS, SVV measurement is invalidated
  - ∘ Normal <200

#### Subcostal (SC)

- Evaluate for pericardial effusion (CLIP)
- Oblique view with IVC (CLIP)
- IVC collapsibility (SAVE)
  - SC oblique, M-mode
  - If variability, measure max and min
- Hepatic vein
  - ∘ IVC view, PW doppler
  - Drains right atrium/IVC, transduces the pressures of the right side
  - Occurs between two QRS complexes
  - S = ventricular systole, actually seeing atrial diastole
  - D = ventricular diastole, actually seeing atrial systole
  - A = atrial kick, causes small reversal in pressure
- Portal vein
  - ∘ IVC view, PW doppler
  - Normal vein, should have continuous flow
  - "Pulsatility" or "to and fro" pattern = volume intolerance

#### **Pulmonary**

- Gen Abdomen preset
- 6 locations upper, mid, and lower on left and right (CLIP x6)
- Evaluation for B lines
- Rating (0) = absent, (1 zone) = scattered, (>2 zones) = diffuse

# Internal Jugular veins (IJ)

- Linear probe, decrease depth to 5cm
- Measure at HOB 0 degrees and then HOB 90 degrees
- · Only need one side

From:

https://ewrobbins.com/ - ewrobbins.com

Permanent link:

https://ewrobbins.com/doku.php?id=resources:clinical\_tools:pocus&rev=1696426623

Last update: 2023/10/04 13:37

